

- ★ READ CAREFULLY THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
- ★ DO NOT DISCARD THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	: 1 - 20 ppm (*)	0.5 - 10 ppm	0.25 - 5ppm	0.2 - 4 ppm
and Sampling Time	: 1.5 minutes	3 minutes	6 minutes	7.5 minutes

(*) Graduations on the detector tube are based on 1 pump stroke.

Number of pump strokes	: 1 (100mL)	2 (200mL)	4 (400mL)	5 (500mL)
Colour Change:	: Yellow → Red			
Detectable Limit:	: 0.05 ppm (5 pump strokes)			
Operating temperature	: 0 - 40 °C (32-104°F) (Temperature correction is necessary.)			
Aspirating Pump:	: Model AP-20, AP-20S, 400B, AP-1, AP-1S or 400A			

⚠ CAUTION

1. DETECTOR TUBE CONTAINS REAGENTS .
2. PRETREAT TUBE CONTAINS REAGENTS.
3. DO NOT TOUCH THESE REAGENTS DIRECTLY ONCE TUBES ARE BROKEN.
4. KEEP THE TUBES OUT OF THE REACH OF CHILDREN.

NOTICE

1. USE ONLY PUMP MODELS AP-20, AP-20S, 400B, AP-1, AP-1S OR 400A. OTHERWISE, CONSIDERABLE ERROR IN INDICATION WILL OCCUR.
2. BEFORE TESTING, CHECK THE ASPIRATING PUMP FOR LEAKS (REFER TO ITEM 8. INSPECTION OF ASPIRATING PUMP) . ANY PUMPS SHOWING SIGNS OF LEAKAGE SHOULD BE CORRECTED BEFORE USE.
3. DO NOT USE THIS TUBE BEYOND THE STATED OPERATING TEMPERATURE RANGE.
4. STORE TUBES IN A REFRIGERATED PLACE (0-10 °C /32-50°F), AND USE BEFORE EXPIRATION DATE PRINTED ON TOP OF THE BOX.
5. PRIOR TO USE, READ CAREFULLY ITEM 9. USER RESPONSIBILITY.
6. READ THE CONCENTRATION IMMEDIATELY AFTER MEASUREMENT.

2. SAMPLING AND MEASUREMENT:




Fig.1

- ① Break both ends of detector tube and pretreat tube, and connect each end of the detector tube and pretreat tube with rubber tube connector as shown in Fig.1.

⚠ CAUTION SAFETY GLASSES AND GLOVES SHOULD BE WORN TO PREVENT INJURY FROM SPLINTERING GLASS.

- ② Insert the detector tube into aspirating pump securely as shown in Fig.1 (Arrow mark shall point to the pump).
- ③ Align the guide marks on the shaft and stopper of the aspirating pump.
- ④ Pull the pump handle at full stroke until it locks and wait for 1.5 minutes or until the completion of sampling is confirmed with the flow indicator of the pump (See descriptions about the flow indicator in the instructions of the pump).
- ⑤ On completion of sampling, read the scale at the maximum point of the stained layer.

REMARKS:The whole reagent is likely to show pale pink discolouration even if Acrylonitrile does not exist. However, the reading concentration which will be affected because the colour can be distinguished from the red discolouration which will be made with Acrylonitrile.

- ⑥ When concentrations are below the scale range, multiple pump strokes can be used to determine these lower concentrations. Use 2, 4 or 5 pump strokes then divide the temperature correction value by each number of pump strokes.

SPECIAL NOTE: I. The scale is calibrated at 20 °C (68°F), 50 %R.H. and 1013hPa. Readings obtained in other circumstances should be corrected (**REFER TO ITEM 3. CORRECTION FOR AMBIENT CONDITIONS**).

II. When the maximum point of the stained layer is unclear or obliquely, read the scale at the centre between the longest and shortest points.

3. CORRECTION FOR AMBIENT CONDITIONS:

① Temperature; Correct the tube reading by following temperature correction table.

Temperature Correction Table					
Tube Readings (ppm)	Corrected Concentration (ppm)				
	0 °C (32°F)	10 °C (50°F)	20 °C (68°F)	30 °C (86°F)	40 °C (104°F)
20	26	22	20	18	17
18	23	20	18	16.5	15.5
16	20.5	18.0	16	15	14
14	18	15.5	14	13	12
12	15.5	13.5	12	11	10.5
10	13	11	10	9.5	9
8	10.5	9	8	7.5	7
6	8	7	6	6	5.5
4	5.5	4.5	4	4	4
2	3	2	2	2	2
1	1.5	1	1	1	1

② Humidity; No correction is necessary.


③ Atmospheric Pressure ;

$$\text{True concentration} = \frac{\text{Temperature corrected concentration} \times 1013}{\text{Atmospheric pressure (in hPa)}}$$

4. INTERFERENCES.

Coexistence of less than 350ppm of Butadinen, 600ppm of Toluene, 800ppm of Hexane or 720ppm of Styrene respectively with Acrylonitrile does not affect the reading value.

5. CHEMICAL REACTION IN THE DETECTOR TUBE:



6. DISPOSAL OF TUBE:

USED TUBES SHOULD BE DISPOSED CAREFULLY ACCORDING TO RELEVANT REGULATIONS, IF ANY.

7. HAZARDOUS AND DANGEROUS PROPERTIES OF ACRYLONITRILE :

TLV-TWA. ◆ : 2 ppm

Explosive range in air : 2.8 - 28 %

◆ Threshold Limit Value established by American Conference of Governmental Industrial Hygienists 2004.

8. INSPECTION OF ASPIRATING PUMP:

Checking for leaks;

① Insert sealed, unbroken detector tube into the pump.

② Align the guide marks on the shaft and stopper of the pump.

③ Pull the handle to full stroke and wait for 1 minute.

④ Unlock the handle and allow it to return slowly into the pump by holding the cylinder and handle securely.

⚠ CAUTION HANDLE WILL TEND TO SNAP BACK INTO THE PUMP QUICKLY.

⑤ If the handle returns completely to the original position, the performance is satisfactory.

Otherwise, refer to maintenance procedure in the instruction manual to correct the leakage.

9. USER RESPONSIBILITY:

It is the sole responsibility of the user of this equipment to ensure that the equipment is operated, maintained, and repaired in strict accordance with these instructions and the instructions provided with each Model AP-20, AP-20S, 400B, AP-1, AP-1S or 400A aspirating pump, and that detector tubes are not used which are either beyond their expiration date or have a colour change different to that stated in the Performance specifications.

The Manufacturer and Manufacturer's Distributors shall not be otherwise liable for any incorrect measurement or any damages, whether damages result from negligence or otherwise.